

## FEATURES :

- 4000 Counts LCD display with white backlit.
- Rugged Mechanical and electrical Design.
- mV range with 1000 MΩ Input impedance.
- Auto ranging.
- Frequency, Capacitance, Duty cycle measurement features.
- 20Amp AC/DC current with overload protection.
- CE/ CAT III 1000V/ IP 54.
- 40 position analog bar graph for trend indication.
- Highly linear readings.
- SELECT RANGE, HOLD, REL, LIGHT, MAX/MIN functions incorporated.

## GENERAL SPECIFICATIONS :

- Dimensions (W X H X D) : 94 X 205 X 36 mm approx.
- Weight : 450g. approx.
- Temp. Coefficient : (0.1% x Specified accuracy) /Deg.C

## ENVIRONMENTAL :

- Operating Temperature : 0°C to 50°C
- Storage Temperature : -20° C to 60°C
- Relative Humidity : 80% RH @ 5°C to 31°C, 50% RH @ 31°C to 40°C non-condensing.

## POWER :

- Power Supply : Battery 9V6F22
- Power Consumption : 8 mA typical.
- Low Battery Indication : 'E' is displayed below 4.5V approx.
- Auto Power OFF : After 15min., Idle Sleep mode consumption 3.5 mA approx.

## OVER LOAD PROTECTION :

- Fuse protection for 'mA' Input Terminal : 0.5A/250V fast blow type Ceramic Fuse.
- Fused Protection for 20A Input Terminal : 20A/250V fast blow type Ceramic Fuse.
- VΩ/→/←/→/← Hz (%) : 440 V DC/AC RMS.

## SAFETY :

- Directives for CE Certification : LVD:2006/95/EC., EMC :2004/104/EC
- Measurement Category : CAT III Reinforced insulation.
- Relevant Standard Specification(s) : EN61010-1:2010, EN61326-1:2006
- IP Rating & Water Protection : IP 54

## ACCURACY :

- Common mode Rejection Ratio > 100dB at DC, 50/60 Hz (1kΩ unbalanced).
- Normal Mode Rejection Ratio > 50dB at DC, 50/60Hz.
- Temp. Coefficient = (0.1% x Specified) accuracy per degree centigrade referred to 25°C.
- Accuracy is specified as ± % of reading + digits and is valid at 25 ± 30°C, ≤ 55% RH.
- It is recommended that calibration equipment used to verify the accuracy of the instrument should be 10 times more accurate.

MOTWANE PRESENTS

**NEW M42**

True Rms Digital Multimeter

ISO 9001 : 2008 Certified

Specially designed in our research laboratory & manufactured under stringent manufacturing processes for Service Industries, SMEs, OEMs and Educational Institutions.

M42 is the most reliable and efficient measuring instrument with CE, CAT III and IP 54 certification along with a calibration certificate traceable to national standards and a standard warranty.



www.motwane.com

**MOTWANE**  
EMPOWERING  
MEASUREMENT

## TECHNICAL SPECIFICATION

### DC VOLTAGE

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
400mV	0.1mV	400.0mV	±(0.5%+3)	1050V DC/AC rms
4V	1mV	4.000V	±(0.5%+3)	
40V	10mV	40.00V	±(0.5%+3)	
400V	100mV	400.0V	±(0.5%+3)	
1000V	1V	1000V	±(0.8%+3)	

Note : Input Impedance 10MΩ approx.

### AC VOLTAGE (50Hz-500Hz) Trms

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
4V	1mV	4.000V	±(1%+5)	1050V DC/AC rms
40V	10mV	40.00V	±(1%+5)	
400V	100mV	400.0V	±(1%+5)	
750V	1V	750V	±(1.2%+8)	

NOTE :

- Input impedance:10 MΩ approx. shunted by 30pF approx.
- 400mV Range will be displayed in manual ranging only (Unspecified Accuracy)

### DC "mV" RANGE

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
40mV	0.01mV	40.00mV	±(1%+5)	440 V DC/ AC rms
400mV	0.1mV	400.00V	±(1%+5)	

Input impedance: ≤1000MΩ approx.

### AC "mV" RANGE (50Hz - 1kHz) Trms

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
40mV	0.01mV	40.00mV	±(1%+5)	440 V DC/ AC rms
400mV	0.1mV	400.00V	±(1%+5)	

### DC CURRENT RANGE

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
40mA	0.01mA	40.00mA	±(1%+5)	0.5A/250V DC/AC fuse protection
400mA	0.1mA	400.0mA	±(1%+5)	
4A	1mA	4.000A	±(1%+5)	20A/250V DC/AC fuse protection
20A	10mA	20.00A	±(1.2%+5)	

### AC CURRENT RANGE (50Hz-500Hz)

Only sine wave (Average)

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
40mA	0.01mA	40.00mA	±(1.2%+5)	0.5A/250V DC/AC fuse protection
400mA	0.1mA	400.00A	±(1.2%+5)	
4A	1mA	4.000A	±(1.2%+5)	20A/250V DC/AC fuse protection
20A	10mA	20.00A	±(1.5%+5)	

### RESISTANCE RANGE

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
400Ω	0.1Ω	400.0Ω	±(0.5% +3)	440V DC/AC rms
4KΩ	1Ω	4.000KΩ	±(0.5% +3)	
40KΩ	10Ω	40.00KΩ	±(0.5% +3)	
400KΩ	100Ω	400.0KΩ	±(0.5% +3)	
4MΩ	1KΩ	4.000MΩ	±(1% +5)	

NOTE :

- Open circuit voltage on 400 Ω to 4 MΩ Ranges is 0.44 VDC approx.
- Open circuit voltage on 40 MΩ Ranges is 0.23 DC approx.

### DIODE TEST

Range	Resolution	Open Circuit Voltage	Test Current
4V	1mV	≤ 2.85VDC approx	0.8mA approx

### CONTINUITY TEST

Range	Resolution	Test Current
400.0Ω	0.1Ω	Meter Beeps at <60Ω

Note : Open Circuit Voltage on continuity range is 0.44VDC approx.

### CAPACITANCE RANGE (Auto Ranging Only)

Range	Resolution	Max. Reading	Accurate (rdg+digit)	Overload Protection
40nF	0.01nF	39.99nF	±(3%+10)	440V DC/AC rms
400nF	0.1nF	399.9nF	±(3%+10)	
4μF	1nF	3.999μF	±(3%+10)	
40μF	10nF	39.99μF	±(3%+10)	
400μF	100nF	399.9μF	±(3%+10)	
4000μF	1μF	3999μF	±(5%+20)	

Note: Settling time on 4000μF range is 30 sec. approx.

### FREQUENCY RANGES (Auto Ranging Only)

Range	Resolution	Max. Reading	Accuracy (rdg+digit)	Overload Protection
10Hz	0.001Hz	9.999Hz	±(0.1%+3)	440V DC/AC rms
100Hz	0.01Hz	99.99Hz	±(0.1%+3)	
1KHz	0.1Hz	999.9Hz	±(0.1%+3)	
10KHz	1Hz	9.999KHz	±(0.1%+3)	
100KHz	10Hz	99.99KHz	±(0.1%+3)	
1000KHz	100Hz	999.9KHz	±(0.1%+3)	
10MHz	1KHz	9.999MHz	±(0.1%+3)	

### DUTY CYCLE MEASUREMENT

Range	Resolution	Accuracy (rdg+digit)	Overload Protection
0.1%-99.9%	0.1%	±(0.5%+3)	440V DC/AC rms

Accuracy is specified at <20 VAC rms

Sensitivity: 10Hz to 100KHz >200mV, 1000KHz to 10MHz >2Vrms

NOTES  
1. The instrument is accompanied with Test & calibration sheet. 2. Test facilities can be provided at the factory with the available test set-ups only. 3. The Company's policy is continuous improvement of its products. We therefore reserve the right of any deviation from illustration or specifications without notice. 4. Stated accuracies are valid from 1/10th of range to FS. 5. Accuracy Specified for temperature range of 25°C ± 5°C & 55% RH ± 10%.

MOTWANE PRESENTS

**NEW M42**

True Rms Digital Multimeter

ISO 9001 : 2008 Certified



**The Motwane Manufacturing Company Pvt. Ltd.**

REGD. OFFICE & FACT: Gyan Baug, Nashik Road, Nashik -422 101. Maharashtra, India. Tel.: 91-253-2463752 / 53. Fax :91-253-2463197 E-mail: sales@motwane.com

#### MUMBAI :

223-224, 2nd Floor, Midas, Sahar Plaza, Andheri Kurla Road, J.B. Nagar, Andheri (E), Mumbai : 400059  
Tel.: 91-22-26058895/ 96/ 97  
Fax: 91-22 26058898  
E-mail : mumbai@motwane.com

#### DELHI :

2 D, 2nd Floor, Vandana Building, 11, Tolstoy Marg, Connaught Place, New Delhi -110 001  
Tel.: 011 - 41522 105  
Telefax : 011 - 41522106  
E-mail : delhi@motwane.com

#### BANGALORE :

# 8/12, 3rd Floor, Honnam Complex, South Cross Road, Basavanagudi, Bangalore-560004.  
Tel. : 080 - 26621186 / 26611323  
E-mail : bangalore@motwane.com

#### KOLKATA :

10, Tarak Dutta Road, Eden Terrace, Near Beck Bagan Mithai, Kolkata 700019.  
Tel. : 033 - 22832436  
E-mail : kolkata@motwane.com

